

The Effect of Maternal Obesity on Sonographic Fetal Weight Estimation

Hossam El Din Hussien, Adel Ali Elboghday, and Dyaa' Yassen Ahmed

Department of Obstetrics and Gynecology – Faculty of Medicine – Al-Azhar University
dovermoon@yahoo.com

Abstract: Aim: To detect the possible effect of maternal obesity on the accuracy of sonographically estimated fetal weight in the third-trimester shortly before labor and to compare the accuracy of the estimation between normal weight, overweight, and class I, class II and class III obese groups. **Methods:** This was a prospective study of singleton pregnancies with sonographic fetal weight estimation prior to scheduled delivery. Women were classified according to current body mass index (BMI) into five categories: normal (BMI 18.5–24.9 kg/m², n = 41), overweight (BMI 25.0–29.9 kg/m², n = 44), obese class I (BMI 30.0–34.9 kg/m², n = 40), obese class II (BMI, 35.0–39.9 kg/m², n = 38) and obese class III (BMI ≥ 40.0 kg/m², n = 35). The estimated fetal weight was compared with the actual birth weight, and the difference between them was recorded as the error. **Results:** shows statistically significant difference between US EFW by (gm) and birth weight by (gm) versus body mass index in obesity class II and III. **Conclusions:** Maternal obesity decreases the accuracy of sonographic fetal weight estimation. Clinicians should be aware of the limitations of sonographic fetal weight estimation, especially in obese patients.

[Hossam El Din Hussien, Adel Ali Elboghday, and Dyaa' Yassen Ahmed. **The Effect of Maternal Obesity on Sonographic Fetal Weight Estimation.** *Nat Sci* 2017;15(9):1-7]. ISSN 1545-0740 (print); ISSN 2375-7167 (online). <http://www.sciencepub.net/nature>. 1. doi:[10.7537/marsnsj150917.01](https://doi.org/10.7537/marsnsj150917.01).

Keywords: body mass index, fetal weight, obesity, ultrasonography.

1. Introduction

Obesity is one of the most serious public health challenges of the 21st century. Obesity has reached epidemic proportions worldwide (Tsigos et al., 2008).

A recent American College of Obstetricians and Gynecologists committee opinion, published in 2013, estimated that at least one-third of pregnant women are obese, and 8% are extremely obese (ACOG; 2013).

The clinical significance of obesity in pregnancy is based on the associated obstetric complications. In addition to obstetric complications caused by maternal obesity, obesity may also impair the visualization of the fetal anatomy and degrade image quality, making it difficult or impossible to obtain adequate images for clinical interpretation. Obese patients with predominant subcutaneous fat will have lower quality images than non-obese patients with minimal subcutaneous fat. Ultrasound imaging of obese patients remains challenging due to the adverse effects of adipose tissue on the propagation of sound waves (Hendler et al., 2004 & Hendler et al., 2005).

The prediction of EFW before delivery during the third trimester plays a pivotal role in obstetric practice, with a major impact on antenatal management. Many important clinical decisions depend upon a precise and accurate assessment of sonographic EFW. For example, overestimation of fetal weight before delivery can lead to unnecessary obstetric interventions. Conversely, underestimation of fetal weight can cause delays in essential obstetric interventions (Aksoy et al., 2015).

Aim of the Work

This study aimed to detect the possible effect of maternal obesity on the accuracy of ultrasound fetal weight estimation during the third trimester shortly before labor.

Patient and Methods

A prospective, comparative study was conducted at Sayed Glal University Hospital. The study population was drawn from consecutive patients who underwent sonographic fetal weight estimation within 7 days of delivery and who fulfilled all of the following **inclusion criteria:**

- Singleton pregnancy.
- Cephalic presentation.
- Pregnant between 37-42 weeks.
- Delivered within one week of fetal weight estimation.
- Proper dating L.M.P or 1st trimester US.
- Intact membranes.

Exclusion criteria:

- Oligohydramnios, anhydramnios.
- Any medical problems (i.e. diabetic, hypertensive, heart disease).
- Placental abnormalities (i.e. placenta previa, ablatio placenta and placental attachment abnormalities).
- Congenital fetal anomalies, hydrops, intrauterine fetal death.
- Utrine fibroids.
- Obstetric emergencies, such as antepartum hemorrhage, eclampsia and acute fetal distress.

One hundred and fifty (150) singleton pregnant women who fulfilled the inclusion criteria were included in the study. All pregnant participants were between 37 and 42 weeks of gestation with a singleton cephalic presentation, and none of the participants had any medical or obstetrical problems.

After providing informed consent, each participant completed an enrolment questionnaire that assessed medical information:

- Maternal age.
- Maternal weight.
- Maternal Height.
- Parity.

Gestational age (Gestational age was calculated based on the last menstrual period and was confirmed in all cases using crown-rump length measured during the first trimester).

Body mass index (BMI) was calculated as the weight in kilograms at the current admission visit divided by the height in meters squared.

The women were classified into five BMI categories based on their current BMI, according to the World Health Organization and National Institutes of Health guidelines: normal weight, BMI 18.5–24.9 kg/m²; overweight, BMI 25.0–29.9 kg/m²; obese class I, BMI 30.0–34.9 kg/m²; obese class II, BMI 35.0–39.9 kg/m²; obese class III, BMI ≥ 40.0 kg/m².

Body mass index was used as a measure of relative maternal size because it correlate with decrease of adiposity in pregnant population and allow comparison of relative maternal size in a large population of women with varying heights.

On presentation to the labor and delivery unit ultrasound scans were performed by the members of the fetal medicine unit of sayedglal university hospital Ultrasound examination was performed transabdominally using MINDRAY DC-3 Ultrasound Machine, using convex abdominal probe with Center Frequency: 3.5 MHz.

The Three measurements of each fetal parameters (BPD, HC, AC and FL) were performed in frozen images of subsequent scans and the means of there values were used for further analysis. the fetal BPD was measured in the standard projection of the fetal head (the maximum diameter of transverse section of the fetal skull at the parietal eminences with the following features: a short midline, the cavum septum pellucidum and the thalami) from the outer edge of the proximal parietal bone to the inner edge of the distal parietal bone. HC was measured in the same plane as BPD, with an ellipse measurement tool from frontal to the occipital part of the outer contour of the skull bone, AC was measured in the standard cross-sectional plane at the level of the stomach and umbilical vein / ductus venosus complex by placing an

ellipse around the outer border of the abdomen. FL measured from the proximal end of the major trochanter to the distal meatphysis.

The fetal biometrics and EFW were calculated using a formula based on the descriptions provided by Hadlock et al. EFW was calculated according to the Hadlock formula: $\log_{10} \text{weight} = 1.335 - 0.0034AC \times FL + 0.0316 \text{ BPD} + 0.0457 AC + 0.1623 FL$ In all cases, the sonographic fetal biometric measurements were performed within 7 days before delivery to eliminate possible impact of duration between ultrasound examination and delivery on the accuracy of the measurements.

All neonates were weighted within 30 minutes of the delivery and infant weight was recorded to the nearest gram.

Because the primary objective was to determine how maternal BMI affect the accuracy of sonographic, the EFW was compared with the actual birth weight (ABW), and the difference between the EFW and the ABW (i.e. simple error) was recorded as the error in grams. The percentage error was defined as: $(\text{EFW} - \text{ABW}) \times 100/\text{ABW}$.

The absolute error was defined as: absolute value of $(\text{EFW} - \text{ABW})$. The mean percentage error represented the sum of the positive (i.e. overestimation) and negative (i.e. underestimation) deviations from ABW.

Results

Table (1): Characteristics distribution of the study group.

Parameters	Analysis [N=150]
Age (years)	18-36 [25.03±4.38]
G.A (wks)	37-42 [38.81±1.47]
Parity	
PG	16 (10.7%)
P1	57 (38%)
P2	39 (26%)
P3	27 (18%)
P4	8 (5.33%)
P5	3 (2%)
BMI [wt/(ht) ²]	
Normal	30 (20%)
Overweight	30 (20%)
Obesity class I	30 (20%)
Obesity class II	30 (20%)
Obesity class III	30 (20%)
BMI [wt/(ht) ²]	18.6-42.3 [31.87±7.01]
Us EFW by (gm)	2750-4690 [3716.26±362.61]
Birth Weight by (gm)	2830-4600 [3744.79±344.72]
Mode of delivery	
CS	67 (44.7%)
NVD	83 (55.3%)

SD: Standard deviations

This table shows Subject demographic and clinical characteristics including age, parity, BMI, EFW, ABW and mode of delivery.

Table (2): Demographic and clinical characteristics of the study group.

Parameters	Normal	Overweight	Obesity class I	Obesity class II	Obesity class III	ANOVA	p-value
Age (years)							
Mean±SD	23.57±4.16	24.33±4.06	25.77±5.06	25.53±3.30	25.93±4.91	1.686	0.156
Range	18-35	19-34	18-36	20-33	18-36		
G.A (wks)							
Mean±SD	39.17±1.64	38.63±1.38	38.80±1.47	38.63±1.40	38.83±1.46	0.657	0.623
Range	37-42	37-42	37-42	37-42	37-42		
Parity							
PG	4(13.3%)	6(20.0%)	3(10.0%)	3(10.0%)	1(3.3%)	4.379	0.357
Multipara	26(86.7%)	24(80.0%)	27(90.0%)	27(90.0%)	29(96.7%)		

This table shows no statistically significant difference between body mass index according to demographic and clinical characteristics.

Table (3): Comparison between body mass index according to US EFW by (gm) and birth weight by (gm).

	Normal	Overweight	Obesity class I	Obesity class II	Obesity class III	ANOVA	p-value
Us EFW by (gm)							
Mean±SD	3448.17±416.91	3542.37±271.43	3711.47±295.12	3863.73±218.36	4015.57±269.15	17.604	<0.001
Range	2750-4300	2980-4050	2890-4120	3410-4349	3410-4690		
Birth Weight by (gm)							
Mean±SD	3478.67±387.66	3576.50±232.92	3747.27±288.97	3920.40±219.07	4131.10±255.38	18.909	<0.001
Range	2830-4220	3040-3980	2890-4090	3457-4400	3470-4600		

This table shows highly statistically significant difference between body mass index and US EFW by (gm) and birth weight by (gm).




Fig. (1): Bar chart between body mass index according to US EFW by (gm) and birth weight by (gm).

Table (4): Comparison between body mass index according to mode of delivery.

Mode of delivery	Normal	Overweight	Obesity class I	Obesity class II	Obesity class III	x2	p-value
CS	9 (30.0%)	10 (33.3%)	13 (43.3%)	17(56.7%)	18(60.0%)	8.793	0.046
NVD	21 (70.0%)	20(66.7%)	17(56.7%)	13(43.3%)	12(40.0%)		
Total	30 (100.0%)	30(100.0%)	30(100.0%)	30(100.0%)	30(100.0%)		

Positive correlation and significant between body mass index classification according to CS.

Table (5): US EFW by (gm) and birth weight by (gm) versus body mass index.

BMI Category	Us EFW by (gm)	Birth Weight by (gm)	t-test	p-value
Normal	3448.17±416.91	3478.67±387.66	-0.910	0.116
Overweight	3542.37±271.43	3576.50±232.92	-0.803	0.234
Obesity class I	3711.47±295.12	3747.27±288.97	-0.933	0.105
Obesity class II	3863.73±218.36	3920.40±219.07	-2.181	0.034
Obesity class III	4015.57±269.15	4131.10±255.38	-3.424	0.021

This table shows statistically significant difference between US EFW by (gm) and birth weight by (gm) versus body mass index in obesity class II and III.




Fig. (2): Bar chart between body mass index according to mode of delivery.




Fig. (3): US EFW by (gm) and birth weight by (gm) versus body mass index.

4. Discussion

In this study, we performed this prospective study to detect the possible effect of maternal obesity on the accuracy of ultrasound fetal weight estimation during the third trimester shortly before labor. The clinical significance of obesity in pregnancy is based on the associated obstetric complications. In addition to obstetric complications caused by maternal obesity, obesity may also impair the visualization of the fetal anatomy and degrade image quality, making it difficult or impossible to obtain adequate images for clinical interpretation. Obese patients with predominant subcutaneous fat will have lower quality images than non-obese patients with minimal subcutaneous fat. Although considerable technical advances in ultrasound technology, such as tissue harmonics and multi-Hertz transducer technology, have been made during the past two decades, ultrasound imaging of obese patients remains challenging due to the adverse effects of adipose tissue on the propagation of sound waves (Hendler et al., 2005).

The detrimental impact maternal obesity has on the accuracy of sonography for detection of anomalies has been reported (Racusin et al., 2012). Thornburg et al. (2009) reported that maternal obesity during pregnancy is associated with major limitations in the ability to evaluate fetal anatomic structures. Fuchs et al. (2013) demonstrated the adverse effects of maternal obesity on genetic sonograms during the first and second trimesters. Goetzinger et al. (2013) examined the sensitivity and specificity for extremes of abnormal fetal growth and found no association with the maternal BMI class.

A clinically important challenge for obstetricians is the impact of maternal obesity on sonographic fetal weight estimation. Horton et al. (2014) investigated the effects of obesity on sonographic estimated fetal weight (EFW) measurement during the third trimester prior to delivery.

Precise EFW measurements are necessary to provide adequate and accurate counseling and obstetric care to pregnant women. The accurate prediction of birth weight prior to delivery is critical and represents a valuable tool for determining further obstetric and delivery management. Inaccurate sonographic EFW (overestimation or underestimation) before delivery adversely affects antenatal management and may lead to inappropriate or untimely interventions and perinatal compromise. For example, overestimation of normal weight fetus before delivery during the third trimester may lead to missed diagnosis of fetal macrosomia and this missed diagnosis can cause unnecessary obstetric interventions, such as cesarean section. The major

clinical concern related to reliance on sonographic fetal birth weight prediction is the inherent inaccuracy associated with the technique (Dudley, 2005).

The prediction of EFW before delivery during the third trimester plays a pivotal role in obstetric practice, with a major impact on antenatal management. Many important clinical decisions depend upon a precise and accurate assessment of sonographic EFW. For example, overestimation of fetal weight before delivery can lead to unnecessary obstetric interventions. Conversely, underestimation of fetal weight can cause delays in essential obstetric interventions. In our study, analysis was confined to 150 singleton pregnancies to detect the possible effect of maternal obesity on the accuracy of ultrasound fetal weight estimation during the third trimester shortly before labor.

Aksoy et al. (2015) investigated the possible effect of maternal obesity on the accuracy of sonographically predicted EFW during the third trimester shortly before the induction of labor. This was a prospective study of singleton pregnancies with sonographic fetal weight estimation prior to scheduled delivery. Women were classified according to current body mass index (BMI) into five categories: normal (BMI 18.5–24.9 kg/m², n = 41), overweight (BMI 25.0–29.9 kg/m², n = 44), obese class I (BMI 30.0–34.9 kg/m², n = 40), obese class II (BMI, 35.0–39.9 kg/m², n = 38) and obese class III (BMI ≥ 40.0 kg/m², n = 35). They observed no statistically significant differences among the five study groups in terms of mean gravidity, parity and gestational age.

This study has demonstrated no statistically significant difference between body mass index according to demographic and clinical characteristics.

In the study done by Aksoy et al. (2015), the demographic and clinical characteristics did not differ between the study groups, except for maternal age, which was 25.19 ± 5.39 years, 26.56 ± 6.31 years, 25.30 ± 5.52 years, 30.42 ± 5.18 years and 30.20 ± 5.88 years in the normal weight, overweight, class I, class II and class III groups, respectively.

In our study, there was a highly statistically significant difference between body mass index and US EFW by (gm) and birth weight by (gm).

Aksoy et al. (2015) observed no significant differences between the groups with respect to EFW and ABW. When intra-group comparisons between EFW and ABW were made, significant differences were found in the obese classes II and III groups. Significant differences in the mean absolute error and the mean absolute percentage error were found between all five groups. A significant difference in the magnitude of the mean absolute error and the absolute

percentage error was observed with increasing maternal obesity.

In our study, a significant positive correlation between body mass index classification according to delivery by CS. Also, there was a statistically significant difference between US EFW by (gm) and birth weight by (gm) versus body mass index in obesity class II and III.

Wolfe et al. (1990) analyzed data from 1622 examinations that were performed at a mean gestational age of 28.5 weeks to determine whether maternal obesity affected visualization of fetal anatomy. They reported a greater risk of suboptimal visualization when BMI (kg/m²) was above the 90th percentile.

Another study conducted by **Dashe et al. (2009)** showed that increasing maternal BMI limits the visualization of the fetal anatomic structures during a standard second-trimester ultrasound examination.

Thornburg (2013) analyzed 112 women who underwent standard ultrasound examination over a 5-year period.

Dammer et al. (2013) have investigated the factors that affect sonographic EFW prediction evaluating the effect of nine different factors, including maternal BMI; presentation of the fetus; time interval between estimation and delivery; fetal gender; fetal weight; placenta location; amniotic fluid index; gestational age and degree of operator experience, on the accuracy of EFW measurements. That retrospective study, which was conducted on 820 singleton pregnancies with gestational age ranging from 22 to 42 weeks, reported that of the nine evaluated factors that may affect accuracy of EFW measurements, only time interval >7 days between estimation and delivery had an adverse effect on prediction.

Caughey (2012) summarized the impact the EFW can have on the mode of delivery. A study by Little et al found that patients who underwent a recent sonographic examination were 50% more likely to undergo a cesarean delivery, with an even greater impact if the EFW was greater than 3500 g. This finding lends credence to the conclusion that clinicians rely on the EFW in their management of labor and decision making regarding the mode of delivery.

Kritzer et al. (2014) quantitated the impact, if any, an increasing maternal BMI has on the accuracy of sonographic EFW obtained within 2 weeks of delivery. Estimation of the EFW near delivery does not appear to be similarly affected by the maternal body habitus. Sonography performed in a dedicated obstetric ultrasound unit within 2 weeks of delivery had a relatively low percentage error for estimation of fetal weight, and this error rate did not vary substantially by maternal BMI classification.

Aksoy et al. (2015) found significantly higher mean absolute error and mean absolute percentage error in the higher BMI category. Strong positive correlations were observed between BMI and the mean absolute error or the mean absolute percentage error; these correlations were statistically significant. Therefore, maternal obesity decreases the accuracy of sonographic fetal weight estimation, in our study there was a statistically significant difference between US EFW by (gm) and birth weight by (gm) versus body mass index in obesity class II and III.

Conclusion and recommendation

It is concluded from this study that obesity brings many health hazards on obese mothers and their babies as obese mothers exposed to cesarean section delivery, adverse pregnancy outcome on their babies as preterm baby, macrosomic baby and congenital anomalies.

Obesity shows strong associations with antenatal complications including increased incidence of pre-eclampsia, gestational hypertension, gestational diabetes and delivery complications including, premature rupture of membrane, preterm delivery, macrosomia, shoulder dystocia, induction of labor, cesarean delivery and postnatal complications including postpartum hemorrhage and postoperative urinary tract infection while underweight women appear to have better pregnancy outcomes than even women with BMI within the normal range.

Even moderate overweight has a significant deleterious effect on the outcome of pregnancy, and obesity leads to major maternal and fetal complications.

Our study has shown that increasing maternal obesity decreases the accuracy of sonographic EFW measurement. Clinicians should be aware of the limitations of sonographic EFW prediction, especially in obese patients.

References

1. Aksoy, H., Aksoy, Ü., Karadağ, Ö. İ., Yücel, B., Aydın, T., & Babayigit, M. A. (2015). Influence of maternal body mass index on sonographic fetal weight estimation prior to scheduled delivery. *Journal of Obstetrics and Gynaecology Research*, 41(10), 1556-1561.
2. American College of Obstetricians and Gynecologists. (2013). ACOG. Practice Bulletin No.137: gestational diabetes mellitus. *Obstet. Gynecol.* 122(2 Pt 1), 406-416.
3. Caughey AB (2012): Obstetrical ultrasound for the estimated fetal weight: is the information more harm than benefit? *Am J Obstet Gynecol*; 207:239-240.

4. Dammer U, Goecke TW, Voigt F et al. Sonographic weight estimation in fetuses with breech presentation. *Arch Gynecol Obstet* 2013; 287: 851–858.
5. Dashe, J. S., McIntire, D. D., & Twickler, D. M. (2009). Maternal obesity limits the ultrasound evaluation of fetal anatomy. *Journal of ultrasound in medicine*, 28(8), 1025-1030.
6. Dudley NJ. A systematic review of the ultrasound estimation of fetal weight. *Ultrasound Obstet Gynecol* 2005; 25: 80–89.
7. Fuchs F, Houllier M, Voulgaropoulos A, et al (2013). Factors affecting feasibility and quality of second-trimester ultrasound scans in obese pregnant women. *Ultrasound Obstet Gynecol* 2013; 41: 40–46.
8. Hendler, I., Blackwell, S. C., Bujold, E., Treadwell, M. C., Wolfe, H. M., Sokol, R. J., & Sorokin, Y. (2004). The impact of maternal obesity on midtrimester sonographic visualization of fetal cardiac and craniospinal structures. *International journal of obesity*, 28(12), 1607-1611.
9. Hendler, I., Blackwell, S. C., Bujold, E., Treadwell, M. C., Mittal, P., Sokol, R. J., & Sorokin, Y. (2005). Suboptimal Second-Trimester Ultrasonographic Visualization of the Fetal Heart in Obese Women Should We Repeat the Examination?. *Journal of ultrasound in medicine*, 24(9), 1205-1209.
10. Horton A, Diaz J, Mastrogiannis D. Accuracy of estimated fetal weight by ultrasonography compared with the Leopold maneuver and effect of maternal obesity. *Obstet Gynecol* 2014; 123 (Suppl): 193S.
11. Goetzinger KR, Tuuli MG, Odibo AO, et al. (2013): Screening for fetal growth disorders by clinical exam in the era of obesity. *J Perinatol*; 33:352–357.
12. Racusin, D., Stevens, B., Campbell, G., & Aagaard, K. M. (2012, June). Obesity and the risk and detection of fetal malformations. In *Seminars in perinatology* (Vol. 36, No. 3, pp. 213-221). WB Saunders.
13. Thornburg, L. L., Miles, K., Ho, M., & Pressman, E. K. (2009). Fetal anatomic evaluation in the overweight and obese gravida. *Ultrasound in Obstetrics & Gynecology*, 33(6), 670-675.
14. Thornburg LL. (2013). Re: Factors affecting feasibility and quality of second-trimester ultrasound scans in obese pregnant women. *Ultrasound Obstet Gynecol*; 41: 7.
15. Kritzer S, Magner K and Warshak CR. Increasing maternal body mass index and the accuracy of sonographic estimation of fetal weight near delivery. *J Ultrasound Med* 2014; 33: 22173-2179.
16. Wolfe HM, Sokol RJ, Martier SM et al. Maternal obesity: A potential source of error in sonographic prenatal diagnosis. *Obstet Gynecol* 1990; 76: 339–342.

7/9/2017